



Running Contract Details	
Equipment Name	HOSPITAL FURNITURE
Running Contract Valid Till	21-03-2024
Tender Ref No	KMSCL/EP/T414/328/2021
Tendered Quantity	200
Supplier Name	M/s Medimek Industries
GST No	27ADBPC1957E1ZY
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
B-50 Additional MIDC Anand Nagar Ambernath(East) - 451506 Thane Maharashtra	Contact Person	Jobin Augustin
	Phone	
	Mobile No	8281815893
	Email	medimek@ymail.com, medimek@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Examination Cot <i>Model & Make : Mi-4011-X</i>	22302 Incl.GST :18%	1561.14	23863.14
		22302	1561.14	23863.14

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Examination Cot

1. Over all approximate size 1890mm Lx 560 mm W x 840 mm H.
2. Fixed upholstered top in two pieces of 56mm width, Rexin shall be of good quality, Mattress shall be U form of density 50.
3. Thickness of sheets shall be 1.2mm. Thickness of frames shall be 2mm.
4. Head rest adjustable on gas spring
5. Upper section of box with three drawers.Dimension of drawers30 Lx50 Wx10 H
6. BP apparatus tray provided near head rest
7. Should provide single foot step
8. Pre-treated and powder coated for foot step.
9. The entire examination table Pre-treated and powder coated.
10. Note : Material test certificate for the proof of SS 304 shall be furnished along with the invoice and the manufacturer shall also submit undertaking that the tested material is used for the construction of the equipment supplied