



Running Contract Details	
Equipment Name	Tonometer Non contact
Running Contract Valid Till	29-11-2024
Tender Ref No	KMSCL/EP/T443/124/2022
Tendered Quantity	20
Supplier Name	M/s Nethra Instrument
GST No	33AAFFN1864Q1ZU
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
MIG 134 - 8th Street Mogappair Eri Scheme Chennai - 600037	Contact Person	Mr.R Vasimalai, Mr. R Anandhan
	Phone	044-26530672
	Mobile No	9840033073, 9847063262
	Email	nethra.ints@gmail.com, nethravaasu@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	TONOMETER NON CONTACT WITH PACHYMETRY <i>Model & Make : HNT-1P/ Huvitz Co. Ltd</i>	666400 Incl.GST :12%	49147	715547
		666400	49147	715547

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
TONOMETER NON CONTACT WITH PACHYMETRY							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehen sive	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :TONOMETER NON CONTACT WITH PACHYMETRY

1. Should have manual and automatic operation.
2. IOP measurement range should have 1mmHg to 60 mmHg
3. Should have minimum 11mm working distance.
4. Should have CCT correction function.
5. Should have fixation light (Blinking and non-blinking)
6. Chin rest vertical movement should have minimum ± 30 mm.
7. Should have minimum ± 22 mm forward /backward movement of measuring unit.
8. Should have minimum horizontal distance is ± 43 mm and vertical distance ± 17 mm of measuring unit.
9. Should have monitor for display.
10. Should have built in printer
11. Should be supplied motorised table with locking wheel.
12. Should be operated from 200 to 240 V/50Hz power supply.
13. Should be supplied 10 Nos paper roll for printer.
14. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid