



Running Contract Details	
Equipment Name	Photo therapy unit single surface LED Type
Running Contract Valid Till	29-07-2023
Tender Ref No	KMSCL/EP/T393/341A/2021
Tendered Quantity	100
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCP4091F1ZS
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Photo therapy unit single surface LED Type <i>Model & Make : BL-50 / Zhengzhou Dison</i>	23576 Incl.GST :12%	1738.73	25314.73
		23576	1738.73	25314.73
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
	Photo therapy unit single surface LED Type			
Labour	1,000.00	1,000.00	1,000.00	
Comprehen sive	2,200.00	2,200.00	2,500.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Photo therapy unit single surface LED Type

photo Therapy Unit Single Surface Led Type

1. Phototherapy should be based on advanced LED technology
2. In LED technology the irradiance should cover the entire treatment area of at least (50 x 25) cms
3. The irradiance should be measured and reported to the user institution at the time of installation and thereafter during every subsequent warranty/CMC/AMC visit and all breakdown visits
4. Irradiance level should be 35 microwatts/cm² at 35 cms or better distance.
5. Lamp source should be continuous tiltable to 45 degrees angle to cover the entire treatment area.
6. System should be height adjustable at least(110-160) cms from floor level with built-in non – resettable timer
7. The unit should have 3 castor wheels with minimum 2 brakes
8. Should work with input 200 to 240Vac 50 Hz supply.
9. Should supplied with 50 numbers eye masks along with the machine
10. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.