



Running Contract Details	
Equipment Name	ECG machine single channel
Running Contract Valid Till	26-07-2025
Tender Ref No	KMSCL/EP/T494(R)/37/2023
Tendered Quantity	300
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	ECG Machine Single Channel <i>Model & Make : CARDIOFAX/ MEDICAID SYSTEMS</i>	24528 Incl.GST :12%	1808.94	26336.94
		24528	1808.94	26336.94

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
ECG Machine Single Channel			
Labour	500.00	500.00	500.00
Comprehensive	2,000.00	2,000.00	2,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :ECG Machine Single Channel

1. Should be single channel.
2. Should have Automatic and Manual mode
3. Should have visual alarm for open lead
4. Should be able to take minimum 30 ECG with a fully charged battery
5. Should have a standard thermal paper (should be available in Local Market)
6. Should have maintenance free digital thermal array printer.
7. Should be compact and portable.
8. Should have ECG lead annotation facility.
9. Should supply the following accessories.
 1. 2 patient cable
 2. eight clip on electrodes (thread type)
 3. 12 chest electrode with silicon rubber bulb
 4. 12 packets of recording paper
 5. bottle of jelly
 6. 12 nos of reusable button type electrode
10. Should operate on mains (220v-50Hz) and rechargeable battery.
11. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid