



Running Contract Details	
Equipment Name	Electrosurgical System (Dental)
Running Contract Valid Till	07-06-2024
Tender Ref No	KMSCL/EP/T416(R)/908/2021
Tendered Quantity	6
Supplier Name	M/s TRITVAM CARE SERVICES
GST No	32AANFT7993G1ZL
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
6/1-A4 TRITVAM TOWERS OPP METRO PILLAR NO 125 AMBATTUKAVU ALUVA-683106	Contact Person	SANDEEP, VINEETH M K
	Phone	0484 2626385
	Mobile No	8891015607,8593005583
	Email	sandeep.nm@erbe-india.com,vineeth.tritvam@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Electrosurgical System (Dental) <i>Model & Make : Vio 100 C</i>	308000 Incl.GST :12%	22715	330715
		308000	22715	330715

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Electrosurgical System (Dental)				
Labour	11,000.00	11,000.00		11,550.00
Comprehensive	5,500.00	5,500.00		5,777.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electrosurgical System (Dental)

Equipment Name: Electrosurgical System (Dental)

1. Should provide state of the art system for precise and efficient tissue contouring
2. Should provide single/multiple control for easy mode selection and intensity adjustment
3. Should provide the basic modes of operation, single/multiple control for easy mode selection
4. Should be provided with waterproof footswitch for hands free electrode activation
5. Should be provided with the following accessories
6. Electrode set should contain at least Coagulation ball – 1 No, 45o straight knife1 No, 45o long loop – 1 no
7. Dispersive electrode
8. Footswitch assembly
9. Hand piece holder
10. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.
11. The maximum power output should be less than 100 watts