

Running Contract Details	
Equipment Name	Electronic Micro Pipette
Running Contract Valid Till	08-04-2023
Tender Ref No	KMSCL/EP/T396/1552/2020
Tendered Quantity	10
Supplier Name	M/s BIOSYS
GST No	32AOEPM7387N1ZC
Installation & Delivery Period	
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
TC 42/2473 VNRA 85 VIVEKANANDA NAGAR EDAPPAZHANJI THIRUVANANTHAPURAM - 695 006	Contact Person	V.MANOSH
	Phone	0471-2352883
	Mobile No	98472 45887
	Email	biosys.tvn@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Electronic Micro Pipette <i>Model & Make : 4987 000.380/Eppendorf</i>	128268.36 Incl.GST :18%	8978.79	137247.15
2	10 ml Pipette Tips(10*96 Pack) <i>Model & Make : 0030 089 464</i>	8894.84 Incl.GST :18%	622.64	9517.48
3	50 ml Pipette Tips (10*96 Pack) <i>Model & Make : 0030 089 480</i>	11512.08 Incl.GST :18%	805.85	12317.93
		148675.28	10407.27	159082.55

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Electronic Micro Pipette							
Labour	5,490.00	5,490.00	5,490.00	5,490.00	5,490.00	5,490.00	5,490.00
Comprehensive	8,784.00	8,784.00	8,784.00	8,784.00	8,784.00	8,784.00	8,784.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electronic Micro Pipette

Electronic Micro Pipette

1. Should be a single channel pipette that can aspirate up to 50 ml of solution
2. Should have volume ranges of 1ul-50 ml
3. Should be able to aspirate viscous solutions also
4. Should have a digital display
5. Should be able to prevent aerosol contamination
6. Should provide with the stand cable and charger
7. Should provide with the adapter for 50 ml tips
8. Should provide 10 numbers each of 10 ml and 50 ml tips for the pipette
9. The rate of 10ml and 50ml pipette tips should be quoted in BOQ for future purchases
10. Yearly calibration should be done at free of cost during the warranty and CAMC period. A Certificate with traceability shall be provided to the user institution.