



Running Contract Details	
Equipment Name	OT Light shadow-less ceiling mount Single dome
Running Contract Valid Till	06-02-2025
Tender Ref No	KMSCL/EP/T455/90/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	<b>OT Light shadow-less ceiling mount Single dome</b> <i>Model &amp; Make : Medik Enterprises</i>	69620 Incl.GST :18%	4873.4	74493.4			
		<b>69620</b>	<b>4873.4</b>	<b>74493.4</b>			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>OT Light shadow-less ceiling mount Single dome</b>							
Labour	600.00	600.00	600.00	600.00	600.00	600.00	600.00
Comprehensive	1,200.00	1,200.00	1,200.00	1,200.00	1,200.00	1,200.00	1,200.00

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :OT Light shadow-less ceiling mount Single dome**

1. Single dome ceiling type Operation Theatre light shadow-less- LED Type
2. Light output should be minimum 1,00,000 lux at 1 meter distance.
3. Should have Flower or Dome Type Light.
4. Number of LED's should be at least 20 and LED shall be replaced free of cost during warranty or CAMC Period.
5. Minimum light field size should be 200mm
6. Power supply to the lamp should be from a constant voltage source/SMPS.
7. Colour temperature should be between 4000Å;K to 5000Å;K
8. Should have short circuit protection for the low voltage power supply
9. Should have user friendly lamp replacement
10. Should have heat reduction filters
11. Should have field focus control through sterile handle. Additional sterile handles two numbers should be supplied
12. Five spare bulbs should be supplied.
13. Minimum spring arm stroke of 500mm.
14. Minimum four step intensity control
15. Should have minimum action radius of 1500mm
16. Input supply – 230Vac, 50Hz.
17. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid
18. Colour rendering index should be 95 or more.