



Running Contract Details	
Equipment Name	OT table manual hydraulic
Running Contract Valid Till	06-05-2023
Tender Ref No	KMSCL/EP/T383/92/2020
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	OT table manual hydraulic <i>Model & Make : MH 1000 Medik Enterprises</i>	146320 Incl.GST :18%	10242.4	156562.4			
		146320	10242.4	156562.4			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
OT table manual hydraulic							
Labour	3,700.00	3,700.00	3,700.00	3,700.00	3,700.00	3,700.00	3,700.00
Comprehensive	9,900.00	9,900.00	9,900.00	9,900.00	9,900.00	9,900.00	9,900.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :OT table manual hydraulic

OT Table Manual Hydraulic

1. The table should have a minimum of 4 sections.
2. The table should have an easily detachable split leg and an easily detachable head section.
3. The tabletop should have a minimum height of 780mm or lesser.
4. The tabletop should have a minimum vertical stroke of 250mm.
5. Should have a sealed hydraulic mechanism to avoid oil spillage.
6. Should have at least 25 degrees Trendelenburg and reverse Trendelenburg
7. Should have at least 20 degrees lateral tilt movement.
8. Should have at least 80 degrees back up movement with the gas spring mechanism.
9. The head section should have up and down movement.
10. The leg section should have 90 degrees down movement and should move sideways to a minimum of 90 degrees.
11. The table should have a heavy and sturdy base and compact to provide adequate foot room for the operating team.
12. The table should be mounted on heavy-duty casters which offer enhanced weight bearing capacity and free mobility.
13. The table should have a single lever foot-operated brake pedal.
14. Should have a minimum patient weight-bearing capacity of 120 Kgs.
15. Base should be made of cast iron and all other parts and accessories should be completely made of Stainless Steel 304 grade except the cushion and gas spring.
16. The table should be supplied with the following accessories.
 - a. Mattress for the complete tabletop in sections - 1 set
 - b. A pair of arm boards with pad and fixing clamp - 1
 - c. A pair of padded shoulder support with clamps - 1
 - d. A pair of padded lateral support with clamps - 1
 - e. A pair of leg crutches with clamps - 1
 - f. Anesthetic screen frame with clamp - 1
 - g. Patient restraint strap - 1.
17. The table should have the facility for easily detachable kidney bridge attachment or an inbuilt kidney bridge.