



(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	Suction low pressure		
Running Contract Valid Till	20-10-2025		
Tender Ref No	KMSCL/EP/T492/120/2023		
Tendered Quantity	300		
Supplier Name	M/s lakshmi Agency		
GST No	33EMMPK9312M1ZL		
Installation & Delivery Period	8 Week(s)		
Up-time / PM vist	95% & 2 Visits per year		
Warranty period	1 Years		

Supplier`s Details					
Address	Contact Details				
CHINNA ARANMANAI STREET	Contact Person	Kudiyarasan			
NEAR KAAMCHI AMMAN KOVIL ARIYALUR -621704	Phone				
	Mobile No	9843457722			
	Email	lakshmiagencytnmsc@gmail.com			

	Item-wise Price Details					
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	Suction low pressure Model & Make : PSP-1/LIFELINE MEDICAL	4928 Incl.GST :12%	363.44	5291.44		
		4928	363.44	5291.44		

Other terms & conditions

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- $2. \ The \ supplier \ shall \ submit \ performance \ security \ amounting \ to \ 5.00\% \ of \ the \ value \ of \ the \ supply \ order.$
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment: Suction low pressure

- 1. Should operate on 50Hz 230 V mains.
- 2. The bottle capacity should be between 100-120ml.
- 3. Should have a vacuum meter and regulation valve.
- 4. Should have a maximum vacuum of 120m bar and should be adjustable.
- 5. Bottle should have air tight lid.
- 6. Should have an on/off switch

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid