# KMSCL

# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details						
Equipment Name	Horizontal Rectangular Autoclave - 400 L					
Running Contract Valid Till	24-04-2025					
Tender Ref No	KMSCL/EP/T276/8C/2018					
Tendered Quantity	50					
Supplier Name	M/s Kerala Surgical Equipment Co					
GST No	32AACFK7295D1ZG					
Installation & Delivery Period	8 Week(s)					
Up-time / PM vist	95% & 4 Visits per year					
Warranty period	3 Years					

Supplier`s Details									
Address		Contact Details							
No.41/1737 A	Contact Person	Mr. Abbas							
MES Centre Near Town Hall Banerji Road	Phone								
Cochin -18	Mobile No	9539012460							
	Email	info@keralasurgicals.com							

				Item-w	ise Price Details				
#	Item Details			Unit Rate (Incl.all taxes & charge	Service C (Through K	_	Grand Total		
1	Horizontal Rectangular Autoclave - 400 L  Model & Make : 60 RS/E / Medipharm				851370 Incl.GST :18%		59595.9	910965.9	
2	Loading Cart				6785 Incl.GST : 1		4749.5	72599.5	
3	Trolley				607' Incl.GST : 1		4253.9	65023.9	
					97999	90	68599.3	1048589.3	
			Annual / C	Comprehensive	Maintenance Cha	arges (Exl.Tax)			
Rate	!	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year	
			Ho	rizontal Recta	ngular Autoclave	- 400 L			
Labour		14,406.00	14,406.00	14,406.00	14,406.00	14,406.00	14,406	.00 14,406.00	
Compi	Comprehensi 44 ve		44,067.00	44,067.00	44,067.00	44,067.00	44,067	.00 44,067.00	

## Other terms & conditions

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- 2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
- 5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

# **Technical Specification**

### Equipment: Horizontal Rectangular Autoclave - 400 L

- 1. Should have triple walled construction of capacity 400 litres or above.
- 2. Should be electrically operated with rectangular chamber shape.
- 3. Chamber dimension should be 600mm \* 600mm \* 1200mm size.
- 4. The Inner chamber (6mm or above thickness), back plate, Door (10mm thickness) should be made up of SS 316 grade.
- 5. The Outer covering & Jacket should be made up of SS 304 grade.
- 6. The Steam generator, heater plate, pipelines, stand should be made up of SS 304 grade.
- 7. Should have water inlet and outlet valves.
- 8. Should have water level gauge
- 9. Should have gauges for measuring inner and outer chamber pressure.
- 10. Should have inner chamber temperature indicator.
- 11. Should have automatic pressure switch and safety valve.
- 12. Should have joint less silicon rubber door gasket.
- 13. Should have provision to adjust the level of the autoclave with respect to the floor.
- 14. Should have vacuum breaker and steam trap valve.
- 15. Should have single piece self locking door which cannot be opened under pressure.
- 16. The complete sterilization cycle should be controlled from one point with the help of a multi-port valve.
- 17. Should have low water protection cut off with audible alarm and should not use high voltage earthing system.
- 18. Should be ISI certified.
- 19. Should be operated on 400 to 440 Vac, 50 Hz.
- 20. Should have 121°C working temperature at 15 psi.

- 21. Heating capacity should be 18 KW or above.
- 22. Should be supplied with internal Tray made up SS.
- 23. Should quote prices of loading cart made of SS 304 & Trolley separately