



Running Contract Details	
Equipment Name	Cycler for performing peritoneal Dialysis (CAPD) Type B
Running Contract Valid Till	11-06-2025
Tender Ref No	KMSCL/EP/T482(R)/1346B/2022
Tendered Quantity	10
Supplier Name	M/s Mitra Industries Pvt. Ltd.
GST No	06AAACM4564C1ZT
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
14/4 Delhi-Mathura Road Faridabad -121003	Contact Person	Mr. Sudhir Kumar ,Ms. Kavita
	Phone	
	Mobile No	8870690006, 7678188931
	Email	sudhir.kumar@mipl.in, kavita@mipl.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Cycler for performing peritoneal Dialysis (CAPD) Type B</b> <i>Model &amp; Make : M- Cycler/Mitra Industries Pvt. Ltd.</i>	486942.4 Incl.GST :12%	35912	522854.4
		<b>486942.4</b>	<b>35912</b>	<b>522854.4</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Cycler for performing peritoneal Dialysis (CAPD) Type B</b>							
Labour	20,000.00	21,000.00	22,050.00	23,152.00	24,310.00	25,525.00	26,802.00
Comprehensive	40,000.00	42,000.00	44,100.00	46,305.00	48,620.00	51,051.00	53,603.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Cycler for performing peritoneal Dialysis (CAPD) Type B**

1. Required Therapy
  - a. CCPD/PD
  - b. Hi-Dose CCPD
  - c. Tidal
  - d. Hi-Dose Tidal
2. Automatic Calculation of Number of Cycles and Dwell time
3. Built in fluid warmer: 35 to 37 Degree C
4. Therapy time : 0 to 48 Hrs
5. Standard and Low fill Modes
  - a. Low fill Mode fill volume : 60 to 1000ml
  - b. Standard Mode fill volume : 200 to 3000 ml
6. Self Correcting, Continuous and system Error Alarms for maximized Patient safety
7. Built in Event Log & Therapy Log
8. Built in Nurses & Service Menu
9. Program Lock
10. Last fill with Same or Different Dextrose
11. Mode of operation : Continuous
12. Volumetric Accuracy reported : Greater of 1% or +/- 10ml
13. Operating Voltage : 220Vac