



Running Contract Details	
Equipment Name	Over bed table
Running Contract Valid Till	23-07-2026
Tender Ref No	KMSCL/EP/T538/551X/2024
Tendered Quantity	3000
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Over bed table <i>Model & Make : Medik-183/ medik Enterprises</i>	7552 Incl.GST :18%	528.64	8080.64
		7552	528.64	8080.64

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Over bed table

Equipment: Over Bed Table

1. Should have height adjustment facility from 850 mm to 1100 mm with the help of operating lever which activates the gas spring to assist the table top to lift.
2. Gas spring should function smoothly with adjustable height and consistent motion during operation.
3. Table top frame shall be designed to hold the top as well as extension works as a handle for the handling of over bed table.
4. Should have anti-scratch, good surface finish ABS Laminated top having dimension 760mm L x 360mm W approximately.
5. Should be mounted on four 5 cm swivel castors