



Running Contract Details	
Equipment Name	Cervical And Lumbar Traction
Running Contract Valid Till	12-10-2024
Tender Ref No	KMSCL/EP/T458/1584/2022
Tendered Quantity	20
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanniyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cervical And Lumbar Traction <i>Model & Make : PHYSIOTRAC - Plus / M/s Bharat Medical Systems</i>	54768 Incl.GST :12%	4039.14	58807.14
		54768	4039.14	58807.14

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Cervical And Lumbar Traction				
Labour	1,000.00	1,000.00	1,000.00	1,000.00
Comprehensive	2,500.00	2,500.00	2,500.00	2,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cervical And Lumbar Traction

1. Should have Continuous & Intermittent Modes
2. Should be used for the treatment of Pelvic, Cervical ,Lumbar and Upper Dorsal regions
3. Should have LED Indications for forward(Hold), Reverse, Rest, Treatment Over & Check/Remote
4. Should have Buzzer indication for treatment over
5. Should have Digital control Circuitry
6. Should have separate traction force selector for CERVICAL & LUMBAR
7. Patient safety switch should be provided
8. Machine should works only when the patient safety switch is connected
9. Traction Table
 - a. Should have facility to fit any traction unit.
 - b. Should be a two section table.
 - c. Should have armpit for counter traction
 - d. Should have facility to adjust the height of the traction unit.
 - e. Should have powder coated mild steel body.
 - f. Should be supplied with soft mattress.
 - g. Should be supplied with foot rest