



Running Contract Details	
Equipment Name	Ultrasound Pachymeter with A-Scan
Running Contract Valid Till	29-11-2024
Tender Ref No	KMSCL/EP/T443/762/2022
Tendered Quantity	10
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Ultrasound Pachymeter with A-Scan</b> <i>Model &amp; Make : Appascan Max with Pachymeter / Appasamy Associates Pvt Ltd</i>	352800 Incl.GST :12%	26019	378819
		<b>352800</b>	<b>26019</b>	<b>378819</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Ultrasound Pachymeter with A-Scan</b>							
Labour	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00
Comprehensive	15,000.00	15,000.00	15,000.00	20,000.00	20,000.00	25,000.00	25,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Ultrasound Pachymeter with A-Scan**

#### **A. A Scan**

1. Should have a probe frequency of 10MHz.
2. Should have an LED internal fixation light.
3. Should have a measurement range of 12 to 46mm or better.
4. Should have built in IOL calculation using 3rd and 4th generation formula including Hoffer Q, Holladay I, SRK II, SRK/T, Haigision.
5. Should have Total Gain control of 0 to 100db or auto gain.
6. Should have eye typesetting for Phakik, dense cataract, silicon filled, Pseudo PMMA, Pseudo Silicone, Pseudo Acrylic, and Aphakic.
7. Should have Report/Image export format PDF, JPEG, PNG, GIF etc.
8. Should have both contact and immersion mode.
9. Should have LCD touch screen display.
10. Should provide external printer/internal printer.
11. Should provide table suitable for equipment and printer.

#### **B. Ultrasound Pachymeter**

1. Should have a probe sample rate of 50MHz or more.
2. Should have measurement range of 300-999u.
3. Should have single and multi-point measurement.

4. The A Scan and Pachymeter should be in one single system and it should have upgradable to B Scan/UBM.

5. The A Scan/Pachy System should have Integrated DICOM, Worklist modality.

6. Should be supplied with 1 hour back up along with the machine.

C. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.