



Running Contract Details	
Equipment Name	Tissue flotation Bath
Running Contract Valid Till	11-09-2024
Tender Ref No	KMSCL/EP/T448/777/2022
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Tissue flotation Bath</b> <i>Model &amp; Make : RTB-02 B &amp; C INDUSTRIES</i>	15863.92 Incl.GST :18%	1110.47	16974.39
		<b>15863.92</b>	<b>1110.47</b>	<b>16974.39</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Tissue flotation Bath**

1. Should have diameter of 200mm and depth of 90mm.
2. Should have double walled construction with glass wool insulation.
3. Both Inner and outer vessel should be made up of good quality Aluminium with powder coating.
4. Should have temperature range up to 90 degree Centigrade.
5. Should have digital display temperature control.
6. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.