



Running Contract Details	
Equipment Name	Ultrasound unit for physiotherapy
Running Contract Valid Till	05-01-2025
Tender Ref No	KMSCL/EP/T463/130/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Ultrasound unit for physiotherapy</b> <i>Model &amp; Make : IEMP 3029 D/ International ELctro Medical Co</i>	21952 Incl.GST :12%	1618.96	23570.96
		<b>21952</b>	<b>1618.96</b>	<b>23570.96</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Ultrasound unit for physiotherapy**

1. A Solid-state lightweight, compact and portable unit with Microprocessor Control.
2. Microprocessor controlled Digital display for output Continuous & Pulse Modes.
3. Four selectable settings in Pulse Mode Digital Timer Audio Alarm for Timer.
4. Voltage AC 220V Probe Transmitter Freq. 1-3 MHz and it should be detachable.
5. Must provide all options in the application of continuous and pulsed ultrasound
6. Should have two treatment heads one with large treatment surface 5sq cm and one with small surface 0.8sq cm
7. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid