



Running Contract Details	
Equipment Name	Crash Cart
Running Contract Valid Till	05-04-2026
Tender Ref No	KMSCL/EP/T518/286/2024
Tendered Quantity	1500
Supplier Name	M/s IMEDFURNS PRIVATE LIMITED
GST No	33AAGCI1214D1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
61/1A SERVICE ROAD PERIYAPANICHERRI CHENNAI -600128	Contact Person	MR.PRADEEP KUMAR
	Phone	
	Mobile No	9500120662
	Email	tenders@imedfurns.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Crash Cart</b> <i>Model &amp; Make : IMED5606K/ IMEDFURNS</i>	19942 Incl.GST :18%	1395.94	21337.94
		<b>19942</b>	<b>1395.94</b>	<b>21337.94</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Crash Cart**

#### **Equipment: Crash Cart (SS)**

1. Overall size shall be more than 900mm L x 500mm W x 1500mm H.
2. The crash cart should be made of 25.4mmx18G Stainless steel tubular frame work.
3. All metal parts shall be SS304 Grade.
4. Should have push handle on one side.
5. Should have S.S. shelves, six colored removable bins & two polystyrene lockable storage units with three drawers each.
6. Facility to carry ECG Monitors, Defibrillators etc on open areas at top centre and bottom shelves.
7. Should have Stainless steel heavy duty drip stand with 4 hooks.
8. Two accessory mounting brackets to mount accessories anywhere without the need of pre-threaded holes.
9. Crash cart should be mounted on 12.5cms dia non-rusting swiveling dual castor wheels. Two having locking arrangement.
10. Oxygen cylinder stand –Stainless steel shall be provided on one side.

#### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail