



Running Contract Details	
Equipment Name	PCR Machine (Conventional)
Running Contract Valid Till	01-02-2025
Tender Ref No	KMSCL/EP/T454/215A/2022
Tendered Quantity	5
Supplier Name	M/s Siva Medical Services
GST No	32ADFFS6671R2ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palathinkara TC - 28/1838 & TC - 28/1838(1) House No. - TENRA 22 Thycaud Thiruvananthapuram-695014	Contact Person	Govind A. S.
	Phone	0471 2338959
	Mobile No	6282644366,8281110354
	Email	ceo@sivamedicals.com,ivdsales@sivamedicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	PCR Machine (Conventional) <i>Model & Make : Prima 96 Plus TM Thermal Cyclor 96 Wells block/ HiMedia Laboratories</i>	275332.94 Incl.GST :18%	19273.31	294606.25
		275332.94	19273.31	294606.25

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
PCR Machine (Conventional)							
Labour	7,200.00	7,200.00	7,200.00	7,200.00	7,200.00	7,200.00	7,200.00
Comprehensive	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00	18,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :PCR Machine (Conventional)

1. Universal Block with following block capacity in a single unit
 - a. 96 x 0.2ml PCR tube
 - b. 1 x 96 well PCR Plate
2. Gradient over columns/rows & Gradient possibility in at least 6 different annealing temperature
3. Heating & Cooling ramp rates should be identical in both Gradient & normal operation.
4. Temperature control (Ramp rate) speed 5°C/s. (3°C/s (heating) and 2°C/s (cooling)