

Running Contract Details	
Equipment Name	Inverted Phase contrast microscope
Running Contract Valid Till	01-02-2026
Tender Ref No	KMSCL/EP/T507(R)/1644/2023
Tendered Quantity	10
Supplier Name	M/s Radical Scientific Equipments Pvt Ltd
GST No	06AACCR8985N1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
9th Milestone Ambala-Jagadhri Road Khudda Kalan Ambala Cantt-133104 (Haryana)	Contact Person	Mr. Vikram Singh Chauhan
	Phone	
	Mobile No	89030997827
	Email	tenders@radicalscientific.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Inverted Phase contrast microscope <i>Model & Make : RTC-7 Code: NX/Radical Scientific Equipments Pvt. Ltd.</i>	336300 Incl.GST :18%	23541	359841
		336300	23541	359841

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Inverted Phase contrast microscope							
Labour	4,275.00	4,275.00	4,275.00	4,275.00	4,275.00	4,275.00	4,275.00
Comprehensive	8,550.00	8,550.00	8,550.00	8,550.00	8,550.00	8,550.00	8,550.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Inverted Phase contrast microscope

1. LED illumination with more than 40,000 hours service life
2. The stage should be large enough to hold a Petri dish
3. Tube should be binocular
4. Phase contrast slider for condenser positions
5. 1 x bright field and 1 x light rings for phase 1 (PH 1)
6. Should be with auto intensity switching between bright field and phase contrast
7. Should have attachment for bright field and phase contrast
8. 10x/20mm eye pieces with rubber eye guards, HC Plan / antifungal coatings
9. 4-fold revolving nosepiece
 - a. HI PLAN 4x/0.10
 - b. HI PLAN 10X/0.22 PH I
 - c. HI PLAN 20 X/0.30 PH I
 - d. HI PLAN I 40 X/0.50 PH I
10. Exchangeable condenser structure
11. Condenser holder step less adjustable, including collector, aperture diaphragm and filter holder with a working distance of 80 mm and N.A 0.3 which allows the use of large flasks without affecting the optical quality
12. Should have coarse and fine focus on both sides, illumination on/off switch
13. Dust cover for stand

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail