



Running Contract Details	
Equipment Name	AB scan
Running Contract Valid Till	03-06-2026
Tender Ref No	KMSCL/EP/T523/1017/2024
Tendered Quantity	15
Supplier Name	M/s Appasamy Associates PVT. LTD
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	AB scan <i>Model & Make : Model: Marvel II AB - scan Model /appasamy associates</i>	1053360 Incl.GST :12%	77685.3	1131045.3
		1053360	77685.3	1131045.3

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
AB scan							
Labour	37,700.00	47,100.00	56,500.00	65,900.00	75,300.00	84,700.00	94,100.00
Comprehensive	75,300.00	84,700.00	94,100.00	1,03,500.00	1,12,900.00	1,22,300.00	1,31,700.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :AB scan

1. A Scan

1. Should have probe frequency of 10MHz.
2. Should have manual or automatic capture.
3. Should have scanning techniques of direct and immersion type.
4. Should have adjustable time varied gain control.
5. Should measure Anterior Chamber Depth, lens thickness, vitreous and axial length
6. Should have IOL formulas like SRK T,SRK II Binkhorst, Holladay, Hoffer Q

1. BScan

1. Should have probe frequency of 10Mhz or more
2. Should have a scan angle of 20 to 50 or more
3. Should capture 50 frame video clips at up to 20 frames per second.
4. Should have data exporting facility in JPEG and PDF format.
5. Should have recording option
6. Both A scan and B scan should be available in a single unit
7. Should have compact design.
8. Should provide printer.
9. Should have footswitch.
10. Should provide suitable UPS with 30 minutes back up.
11. Should work in 220-240Vac 50 Hz input supply
12. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate/ ICMED Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid