



Running Contract Details	
Equipment Name	Optical Coherence Tomography Model B
Running Contract Valid Till	29-09-2024
Tender Ref No	KMSCL/EP/T451/607B/2022
Tendered Quantity	12
Supplier Name	M/s Appasamy Associates PVT. LTD
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Optical Coherence Tomography Model B <i>Model & Make : Appa OCT 2018/ Appasarry Associates Pvt. Ltd</i>	1792000 Incl.GST :12%	132160	1924160
		1792000	132160	1924160

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Optical Coherence Tomography Model B							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehensive	96,000.00	1,12,000.00	1,28,000.00	1,44,000.00	1,60,000.00	1,76,000.00	1,92,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Optical Coherence Tomography Model B

1	Scan speed	130000 A scan lines per second
2	Vertical scan resolution	15 microns
3	Horizontal scan resolution	5 microns
4	Scan pattern	HD Macula line, Macula 3D, Disc 3D, GCC, Anterior segment
5	Source	SLD for OCT&IR Source for funded imaging
6	Scan depth	2mm in retinal tissue
7	Internal & External Fixation	Available
8	Min pupil diameter required	3.5mm
9	Field angle	30degree
10	Viewing method	22" color TFT monitor for live fundus/OCT imaging
11	Power supply capability	110V/ 220V AC
12	USB printer interface	Should be available
13	Anterior segment OCT imaging should be available	