



| Running Contract Details       |   |
|--------------------------------|---|
| Equipment Name                 | Patient Transport Trolley with Portable Oxygen Cylinder |
| Running Contract Valid Till    | 25-08-2025  |
| Tender Ref No                  | KMSCL/EP/T494(R)/1465/2023                              |
| Tendered Quantity              | 500   |
| Supplier Name                  | M/s Medifocus India Pvt Limited                         |
| GST No                         | 33AAECM6662R1ZU   |
| Installation & Delivery Period | 8 Week(s)   |
| Up-time / PM vist              | 95% & 2 Visits per year                                 |
| Warranty period                | 2 Years   |

| Supplier`s Details  |                 |                         |
|---|-----------------|-------------------------|
| Address   | Contact Details |                         |
| Shop No 1 Portion A Ground Floor<br>Old No 85 New No 24<br>Arya Gowda Road West Mambalam<br>Chennai | Contact Person  | N Sivaramakrishnan      |
|   | Phone           | 42668685                |
|   | Mobile No       | 9841048495              |
|   | Email           | info@medifocusindia.com |

| Item-wise Price Details |  |   |                                    |                 |
|-------------------------|--|---|------------------------------------|-----------------|
| #                       | Item Details   | Unit Rate<br>(Incl.all taxes & charges) | Service Charges<br>(Through KMSCL) | Grand Total     |
| 1                       | <b>Patient Transport Trolley with Portable Oxygen Cylinder</b><br><i>Model &amp; Make : IMED5602C1/Med Furns</i> | 50417.86<br>Incl.GST :18%               | 3529.25                            | 53947.11        |
|                         |  | <b>50417.86</b>                         | <b>3529.25</b>                     | <b>53947.11</b> |

| Annual / Comprehensive Maintenance Charges (Exl.Tax)           |                      |                      |                      |        |
|--|----------------------|----------------------|----------------------|--------|
| Rate   | 3 <sup>rd</sup> Year | 4 <sup>th</sup> Year | 5 <sup>th</sup> Year |        |
| <b>Patient Transport Trolley with Portable Oxygen Cylinder</b> |                      |                      |                      |        |
| Labour   | 378.00               |                      | 378.00               | 378.00 |
| Comprehensive  | 756.00               |                      | 756.00               | 756.00 |

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Patient Transport Trolley with Portable Oxygen Cylinder**

#### **Equipment: Patient Transport Trolley with Portable Oxygen Cylinder**

1. Should have 3 sectional mattress base with chromed or epoxy-coated steel frame
2. Should have an anti static mattress
3. Should have hydraulic height adjustment of 610 mm-980mm
4. Should have anti static castors with braking facility
5. Should have facilities for trendelenburg & reverse trendelenburg positions
6. Should have weight bearing capacity of at least 125 kg
7. Should have place for fixing oxygen cylinder
8. Should have detachable telescopic IV rods with four hooks and foldable side rails
9. Should be supplied with standard accessories such as patient restraining straps – 1set (at least 2 Nos)
10. Patient restraining straps- 1 set (at least 2 nos)
11. Should have side rails