### KERALA MEDICAL SERVICES CORPORATION LTD



(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	RH Viewing Box		
Running Contract Valid Till	15-02-2025		
Tender Ref No	KMSCL/EP/T469(R)/422/2022		
Tendered Quantity	15		
Supplier Name	M/s Labline Instruments		
GST No	32AABFL5818N1Z6		
Installation & Delivery Period	6 Week(s)		
Up-time / PM vist	95% & 0 Visits per year		
Warranty period	1 Years		

Supplier`s Details						
Address	Contact Details					
P B No. 36	Contact Person	Bini				
Eroor Road Trippunithura-682301	Phone	04842776582				
	Mobile No	9020304422				
	Email	lablinecochin@gmail.com				

	Item-wise Price Details						
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	RH Viewing Box  Model & Make : Model LRV I/ Labline Instruments	21240 Incl.GST :18%	1486.8	22726.8			
		21240	1486.8	22726.8			

### **Other terms & conditions**

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- $2. \ The \ supplier \ shall \ submit \ performance \ security \ amounting \ to \ 5.00\% \ of \ the \ value \ of \ the \ supply \ order.$
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

# **Technical Specification**

## **Equipment : RH Viewing Box**

- 1. Suitable for viewing RH Slides
- 2. Should have glare free light provide uniform illumination of slide contents
- 3. Should have temperature control
- 4. Should work in 240V 50Hz supply.
- 5. The product should be CE Certified