



Running Contract Details	
Equipment Name	Shoulder CPM
Running Contract Valid Till	11-09-2025
Tender Ref No	KMSCL/EP/T494/24C/2023
Tendered Quantity	12
Supplier Name	M/s Cnergy Incorporation
GST No	33AAJFC9598G1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
5277 First floor Ram nagar north extension 8th street Madipakkam Chennai-600091	Contact Person	Kumaravel R
	Phone	044-40067976
	Mobile No	9848816400
	Email	tender@cnergy.in

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Shoulder CPM <i>Model & Make : Artromot-S3/DJO Global</i>	615000 Incl.GST :12%	45356.25	660356.25			
		615000	45356.25	660356.25			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Shoulder CPM							
Labour	4,237.28	4,237.28	4,237.28	4,237.28	4,237.28	4,237.28	4,237.28
Comprehensive	13,983.05	16,313.55	18,644.07	20,974.58	23,305.08	25,635.59	27,966.10

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Shoulder CPM

1. Should provide flexion-extension, abduction- adduction and Rotations of shoulder
2. Movements in complete range > 180
3. Movements in different modes – isolation/ oscillation/ stretching etc
4. Range programmable
5. Accessories easily detachable
6. Display with touch screen
7. Timer adjustable program
8. Programmable for individual patients
9. Hand held emergency stop switch
10. Maximum load on motion element >12kg
11. Height adjustable electronically
12. 220-240v 50-60hz
13. US FDA / CE approved