



Running Contract Details	
Equipment Name	Weighing Machine Baby (Electronic)
Running Contract Valid Till	11-02-2023
Tender Ref No	KMSCL/EP/T365(R)/145/2020
Tendered Quantity	200
Supplier Name	M/s South india industries
GST No	32ADCPG7506Q1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
south india industries 43/2690 red cross roadNear CH Fly Over Calicut 673001	Contact Person	Nigilesh
	Phone	04952369477
	Mobile No	9846088820
	Email	info@deltaindia.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Weighing Machine Baby (Electronic) <i>Model & Make : SII-T-B/Delta South India Industries</i>	6947.84 Incl.GST :18%	486.35	7434.19
		6947.84	486.35	7434.19

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Weighing Machine Baby (Electronic)

1. Should have 5 gm accuracy
2. Should have a minimum measuring capacity 25 gms.
3. Should have easy to read LED display
4. Should measure a maximum weight of 20kgs.
5. Should have zero calibration.
6. Should hold the measured value irrespective of the baby movements.
7. Should have electronic damping facility for eliminating the reading fluctuations caused by moving baby.
8. Weighing pan should be suitable for weighing new born babies and the construction should not allow the baby to slip from the tray.
9. The Tray should be made of fiber glass/acrylic.
10. Should have a pan size of at least 50cms length, 20cms width and 8cms height.
11. The pan should have facility to measure the length of baby.
12. Should operate on input supply 200 to 230 Vac, 50 Hz.
13. Should bear certification for each machine by department of legal metrology valid in Govt. of Kerala. The tenderer shall submit copy of the model approval issued by Govt. of Kerala, Legal Metrology department along with technical bid