



Running Contract Details	
Equipment Name	Intense Pulsed Multi Light (IPL) Laser with Radio Frequency
Running Contract Valid Till	20-12-2024
Tender Ref No	KMSCL/EP/T446/1594/2022
Tendered Quantity	5
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Intense Pulsed Multi Light (IPL) Laser with Radio Frequency <i>Model & Make : Platform/ COSMOAESTHETIC A MEDICO PVT LTD</i>	383500 Incl.GST :18%	26845	410345
		383500	26845	410345

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Intense Pulsed Multi Light (IPL) Laser with Radio Frequency							
Labour	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00
Comprehensive	16,250.00	16,250.00	16,250.00	16,250.00	16,250.00	16,250.00	16,250.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Intense Pulsed Multi Light (IPL) Laser with Radio Frequency

1. 4 in 1 IPL, EPL, RF, Q Switched ND Yag Laser along with 3 additional Doctor Protective Goggles and appropriate stabilizer should also be provided
2. Should work on AC 110-240 V 50/ 60Hz
3. Should have minimum 12 inches Color LCD touch screen interface system
4. Should have light protective hand piece.
5. Integrated cooling system by water, air and semiconductor cooling.
6. E- light
 - a. Energy density - IPL: 10-50 J/ cm³ & RF- 10-50 J/cm
 - b. Spot size: 15' 50 mm or 12' 35 mm²
 - c. Discharge interval 1S, 2S, 3S, 4S.
 - d. Probe temp - 3⁰C – 2⁰C
 - e. Wavelength: 640 nm/530 nm/480 nm.
7. RF
 - a. Handle size- 43 mm/ 30 mm/15 mm
 - b. Frequency - 3 p/s.
 - c. Energy - 1-20 J/ cm²
 - d. Probe temp - 2⁰C-0⁰C
8. Laser
 - a. Frequency - 1-10 p/s
 - b. Energy - 10-2000
 - c. Wavelength - 1064/532