



Running Contract Details	
Equipment Name	Ureteroscope (6 to 7.8 Fr)
Running Contract Valid Till	27-03-2025
Tender Ref No	KMSCL/EP/T464/1125/2022
Tendered Quantity	20
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
2nd Floor 12/307-C1 KSIE Building CIC Freight Station ICTT Road Anavathil Jn. Udyogamandal Eloor Ernakulam - 683501	Contact Person	Mr. Viji Vijayan
	Phone	0484-2988006
	Mobile No	99466 02090
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ureteroscope (6 to 7.8 Fr) <i>Model & Make : Karl Storz Ureteroscope (6 to 7.8 Fr) / Karl Storz GmbH & Co Germany</i>	671439.57 Incl.GST :12%	49518.67	720958.24
		671439.57	49518.67	720958.24

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ureteroscope (6 to 7.8 Fr)							
Labour	8,536.00	8,536.00	8,536.00	8,536.00	8,536.00	8,536.00	8,536.00
Comprehensive	28,454.00	29,877.00	31,371.00	32,939.00	34,586.00	36,315.00	38,131.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ureteroscope (6 to 7.8 Fr)

1. 6 to 7.8 Fr
2. Length – 430 to 450 mm
3. Working channel- 4.2 Fr Channel or more
4. 5-7degree Direction of view, angled ocular,semirigid
5. ACCESSORIES
 - a. Attachment ,two –way ,with two irrigation ports
 - b. Sealing cap,0.7 mm hole ,10pcs.,black
 - c. Stone grasping Forceps- 3 to 4 fr,60cm-2
 - d. Biopsy Forceps-4 fr ,60 cm-2
 - e. Stone basket-5
 - f. Brush for cytology-2
 - g. Stainless steel Autoclavable instrument tray-1