

Running Contract Details	
Equipment Name	Urea Breath Analyzer
Running Contract Valid Till	19-01-2025
Tender Ref No	KMSCL/EP/T460(R)/1602/2022
Tendered Quantity	5
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Urea Breath Analyzer <i>Model & Make : Helic ABT Reader/ AMA Co. Ltd</i>	824820 Incl.GST :18%	57737.4	882557.4			
		824820	57737.4	882557.4			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Urea Breath Analyzer							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehensive	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Urea Breath Analyzer

1. Quick result after 15 minutes
2. Automatic testing device for automatic air sampling and reading the color of the indicator tube
3. Automatic testing device does not require periodical maintenance. It doesn't have built in sensitive element
4. Plastic tube sealed from both sides with an indicator composition of a certain length inside (indicator tube)
5. Disposable mouth piece
6. The QR code, printed on a card, is included into every batch of indicator tubes. The code contains information regarding the serial number, production date and shelf-life, and also contains information about various parameters, needed for the device in order to make precise reading of indicator tubes and to provide adequate testing
7. Software for reader, installed on a PC
8. Provides a high sensitivity (95%) and specificity (97%)
9. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid