



Running Contract Details	
Equipment Name	Multi Stainer Workstation
Running Contract Valid Till	25-10-2024
Tender Ref No	KMSCL/EP/T458/1275B/2022
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Multi Stainer Workstation <i>Model & Make : SS 30-H Automatic Slide Stainer with Drying Station / ESPECIALIDADES MEDICAS MYR SL</i>	1869120 Incl.GST :18%	130838.4	1999958.4
		1869120	130838.4	1999958.4

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Multi Stainer Workstation							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehensive	25,000.00	30,000.00	35,000.00	40,000.00	45,000.00	50,000.00	55,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Multi Stainer Workstation

1. Programmable fully automatic walk away stainer will be used primarily for routine staining procedures in histology and cytology laboratories
2. Capable of performing multiple staining procedures randomly or in batch mode.
3. Should have capability of processing upto 1 to 330 slides or more.
4. Should facilitate loading 1 to 20 slides per batch or higher.
5. Should have reagent reservoirs, 160ml each with cover and arranged in two levels
6. Programmable wash reservoirs of two to five with user defined option
7. Should have drying station with paraffin waste tray with temperature range 35-65°C.
8. Should have facility of programming start station and end station.
9. Facility of upto 15 or more programs should be available.
10. Should operate with robotic / similar arm movement and enhanced wash feature.
11. Aspirating fan drawing fumes through filter, optional duct connection adaptor, to be available if required in future
12. Ability to process in monitor basket position.
13. Alarms should be audible beep and screen display messages should be available for user convenience.
14. Diagnostic function like error messages and codes should be available.
15. Safety features like door and cover closure sensors should be available.
16. Should have a provision to connect automated cover slipper in future
17. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission.
18. No of reagent reservoir – 18 or more
19. Drying station - minimum 1