



Running Contract Details	
Equipment Name	Paediatric Fibre Optic Video Bronchoscope
Running Contract Valid Till	18-08-2024
Tender Ref No	KMSCL/EP/T435/1051/2021
Tendered Quantity	5
Supplier Name	M/s St. Luke Med Devices Pvt Ltd
GST No	32AAMCS6936J1ZX
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
ISHI LAND #18Tagore Nagar 1st Street Ayanavaram Chennai-600023.	Contact Person	Mr. Abdul Kadhar
	Phone	
	Mobile No	8589997202, 9710450090
	Email	st.lukecochin@gmail.com, st.lukevision@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Paediatric Fibre Optic Video Bronchoscope <i>Model & Make : FB 8V/Pentax Medical</i>	754970.16 Incl.GST :12%	55679.05	810649.21
		754970.16	55679.05	810649.21

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Paediatric Fibre Optic Video Bronchoscope							
Labour	12,150.00	12,150.00	12,150.00	12,150.00	12,150.00	12,150.00	12,150.00
Comprehensive	24,300.00	24,300.00	24,300.00	24,300.00	24,300.00	24,300.00	24,300.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Paediatric Fibre Optic Video Bronchoscope

I. Pediatric Intubating fiberscope

- a. Pediatric intubating fiberscope OD 2.8 mm (OD 2.8mm for ETT tube 3.5 till 7.50)
- b. Fiberscope for difficult intubation having OD 2.8 mm in order to be compatible for minimum ETT size 3.5 and above with working length strictly not less than 65cm.
- c. Up/ down tip deflection should be same 140/130 degree.
- d. Scope should have length calculation marks preferably 5cm each.
- e. Should have light source connection for LED operated light as well as via light cable for electricity operated light source.
- f. Total length should be ranging between 98cm.
- g. It should have inbuilt suction port which can be attached to already available suction machine and should have an instrument of 1.2mm to use flexible forceps or for oxygenation.
- h. Scope should have very good image quality and can be attached through camera for visualizing on monitor.
- i. Scope should have 90 degree angle of view and 0 degree direction of view.

II. The following accessories to be provided

1. Adaptor to hold the endotracheal tube while intubation
2. Case for flexible scope
3. Test for checking the leakage
4. Pressure compensation cap
5. Disposable suction caps 20
6. Cleaning brush
7. Suction adapters
8. Sterile flexible cover
9. Airway to guide the tube orally
10. Tube holder should be a part of standard accessory
11. Set should include- suction adaptors (Disposable), irrigation adaptors (Disposable), cleaning brush & leakage tester as standard accessories

III. Suitable for following applications

1. Bronchoscopy
2. Endotracheal intubation (Gold standard for difficult airways)
3. Foreign body removal
4. Bronchial lavage
5. Inspection of the airways
6. Dilatation tracheotomy
7. DLT placement for 29 french and more