



Running Contract Details	
Equipment Name	Oto Acoustic Emission Screeners
Running Contract Valid Till	21-10-2026
Tender Ref No	KMSCL/EP/T548/513/2024
Tendered Quantity	90
Supplier Name	M/s Alliance Biomedica Pvt. Ltd
GST No	33AACCA4937D1Z3
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Flat No. 15 AshirwadÃç New No. 30 (Old No. 12) Puliyur First Lane Trustpuram. TamilNadu Chennai 600024	Contact Person	J. KUMAR
	Phone	044 24803704 / 23725299
	Mobile No	09840215864
	Email	jkumar@alliancebiomedica.com,info@alliancebiomedica.com,sreejith.vk@alliancebiomedica.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Oto Acoustic Emission Screeners <i>Model & Make : GSI CORTI (DPOAE & TEOAE) SCREENING UNIT/ Grason Stadier Inc.(GSI), U.S.A .</i>	199999.8 Incl.GST :5%	15733.32	215733.12
		199999.8	15733.32	215733.12

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Oto Acoustic Emission Screeners							
Labour	7,500.00	7,875.00	8,269.00	8,682.00	9,116.00	9,572.00	10,051.00
Comprehensive	30,000.00	31,500.00	33,075.00	34,729.00	36,465.00	38,288.00	40,202.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Oto Acoustic Emission Screeners

Technical Specification

Equipment: Oto Acoustic Emission Screeners

1. Should be hand held and portable
2. TEOAE should be up to 4 Khz and DPOAE upto 5 KHz
3. Should be fully automated and require no interpretation.
4. Should have 2 fixed protocols
5. Should have Thermal Printer
6. Should have Wireless printing
7. Should provide Data Management Software compatible for PC & Tab
8. Should store minimum 250 screening results
9. Battery should have rechargeable type and supplied with a suitable charger
10. Should have Probe tips assorted type
11. Should supply thermal paper rolls (15 nos.) and suitable carry case along with the unit.