



Running Contract Details	
Equipment Name	Biosafety cabinet 3 feet
Running Contract Valid Till	31-12-2024
Tender Ref No	KMSCL/EP/T404/171A/2021
Tendered Quantity	205
Supplier Name	M/s Biolinx India Private Limited
GST No	27AAECB9896Q1ZI
Installation & Delivery Period	12 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
209 Lodha Supremus Road No.22 Wagle Estate Near New Passport office Thane- 400604	Contact Person	NIKHILESH K
	Phone	
	Mobile No	9947662077,9321141976
	Email	nikhileshk20@gmail.com, biolinx@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Biosafety cabinet 3 feet</b> <i>Model &amp; Make : HF Safe 900 LC</i>	388810 Incl.GST :18%	27216.7	416026.7
2	<b>Cost of cancopy</b>	17700 Incl.GST :18%	1239	18939
		<b>406510</b>	<b>28455.7</b>	<b>434965.7</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Biosafety cabinet 3 feet</b>							
Labour	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00
Comprehen sive	7,900.00	7,900.00	7,900.00	7,900.00	7,900.00	7,900.00	7,900.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Biosafety cabinet 3 feet**

1. Microprocessor controlled, Class II Bio safety Cabinet (Type A2)
2. Floor standing model with castor wheel and lock
3. Should have 2, 3, 4 feet wide, 2 ft deep easily removable SS 304 Worksurface
4. Mini pleat ULPA filter with particle retention better than 99.999% for 0.12 micron or above particles
5. All contaminated air paths should be in a negative pressure or covered with a negative pressure plenum.
6. The vertical sliding front door of tempered glass with continuously variable height adjustment.
7. 2 x 40w fluorescent light illumination and 1 x 3ft UV.
8. Body M.S with sufficient protective coating.
9. Air downflow velocity: As per standards
10. Standard inflow velocity: As per standards
11. Work access opening min 8 inches, 10 inches preferable.
12. Integral hour meter to monitor usage.
13. Touchpad switches for ease of use and cleaning.
14. Temperature-compensated, airflow velocity sensors for a dynamic display of inflow and downflow velocity.
15. Automatic dynamic compensation of airflow in the event of any blockage or filter obstruction.
16. Collection tray for spills below the work surface.
17. All transparent windows should be of tempered glass
18. Gas tap - one towards the back of the cabinet.
19. Minimum 2 nos. 5A electrical outlets should be provided inside the workspace.
20. Certification of filter integrity and proper airflow should be done after installation and traceable calibration certificate provided.
21. Timer and interlocking of the front sash for UV light should be available
22. Alarm for any out of range parameters
23. The equipment should be certified as per EN 12469 standards by an external authorized certifying agency.
24. The equipment should be calibrated once a year or sooner if required, and a certificate provided during Warranty and CMC period.
25. Cost of canopy suitable for 2, 3 and 4 feet should be quoted separately which will not be taken evaluation.