



Running Contract Details	
Equipment Name	Emergency Trolley (Trauma Emergency Cot)
Running Contract Valid Till	05-04-2026
Tender Ref No	KMSCL/EP/T522/551Z/2024
Tendered Quantity	500
Supplier Name	M/s Carevel Medical Systems Pvt Ltd
GST No	09AABCC6250C1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
S-35-36 Site- IV Suhibabad Industrial Area Ghaziabad - 201 010	Contact Person	Mr. Gyanendra Rajvanshi
	Phone	0120-4292637
	Mobile No	9599679916
	Email	rkurup@carevelglobal.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Emergency Trolley (Trauma Emergency Cot)</b> <i>Model &amp; Make : C-2002-H /M/s Carevel Medical system Private Limited</i>	45312 Incl.GST :18%	3171.84	48483.84
		<b>45312</b>	<b>3171.84</b>	<b>48483.84</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Emergency Trolley (Trauma Emergency Cot)**

1. Should have two sectional top Xray translucent high pressure laminate with facility to insert Xray cassette from both sides and ends of trolley
2. Should be able to X ray the patient from positions the entire length and width of the trolley
3. Should have ratchet adjusted back section
4. Should have trendelenberg and reverse trendelenberg positions.
5. Should have hydraulic height adjustment.
6. Should have removable stretcher for easy cleaning and disinfections of the Xray platform.
7. Should have place for fixing oxygen cylinder.
8. Should have detachable SS telescopic IV rods and collapsible SS side rails.
9. Body should be pre-treated and powder coated.
10. Should be supplied with standard accessories such as
  - a. SS Collapsible Side rails -1no.
  - b. Height adjustable IV rod-1no.
11. Dimension:-
  - i. Max.Length:1900±20mm
  - ii. Max.Width:700±10mm
  - iii. Height:Min 65±5cm Max90±5cm
  - iv. Trendelenberg:10±2deg.
  - v. Reverse trendelenberg:5±1deg
  - vi. Xray viewing area Entire length
12. The manufacturer should be compliant with ISO9001, 14001, 13485 & OHSAS 18001.
13. Should be provided with mattress of 32 density foam 2" thick with two way stretchable rexine

### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail