



Running Contract Details	
Equipment Name	Slit Lamp with Imaging Facility
Running Contract Valid Till	12-10-2025
Tender Ref No	KMSCL/EP/T489(R)/110G/2023
Tendered Quantity	5
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Slit Lamp with Imaging Facility <i>Model & Make : SL- CAM ICM-7/Appasamy Associate Pvt. Ltd.</i>	728000 Incl.GST :12%	53690	781690
		728000	53690	781690

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Slit Lamp with Imaging Facility							
Labour	19,500.00	19,500.00	19,500.00	19,500.00	19,500.00	19,500.00	19,500.00
Comprehensive	52,000.00	52,000.00	58,500.00	58,500.00	65,000.00	71,500.00	78,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Slit Lamp with Imaging Facility

Equipment: Slit Lamp with Imaging Facility

1. System should have microscope unit with following specification
2. Should have magnification change with 3/5 step drum rotation
3. Should have eyepiece 12.5x
4. Should have magnification steps 10x, 16x, 25x
5. Should have field of view 27mm, 16mm, 11mm
6. Should have pupillary distance 50-80
7. Should have dioptr adjustment range -8D to +8D
8. System should have illumination unit with following specification
 - a. Should have slit width 0-14mm continuous, slit length 0-14 mm continuous, slit apertures ϕ 14, 10, 6, 4, 3, 1, 0.2mm, slit angles- 0° -180°, slit inclination 5°, 10°, 15°, 20°, filter- heat absorbing, green & cobalt blue, illumination type 3.4v, 1W
9. Should have longitudinal in/out 99mm, lateral (left/right) 118mm, vertical (up/down) 30 mm, chin rest range 117mm
10. System should have camera
11. Should have image sensor 12 mega pixel, photo resolution 4056x3040 – 3840 x 2748, format JPEG, frame 30fps, video format H264, video format USB 3.0
12. Should be supplied with PC along with software.

PC Configuration :

- a. Processor : Intel i5 or more (10th generation)
- b. RAM : 8 GB
- c. Hard disk : 1 TB
- d. Monitor : 14 inch or more

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid