



Running Contract Details	
Equipment Name	Apheresis Machine
Running Contract Valid Till	07-05-2026
Tender Ref No	KMSCL/EP/T524(R)/638/2024
Tendered Quantity	5
Supplier Name	M/s Fresenius Kabi India Pvt Ltd
GST No	
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
A-3MIDC Ranjangaon Ganapati Taluka Shirur Pune Dist- 412220	Contact Person	
	Phone	
	Mobile No	
	Email	govindram.cheyyur@fresenius-kabi.com,anand.s@fresenius-kabi.com,rishi.soni@fresenius-kabi.com,ashutoshk.singh@fresenius-kabi.com,vaisak.chandran@fresenius-kabi.com,s.rajagopal@fresenius-kabi.com,sivaprasad.b@fmc-asia.com,guna.fmc@gmail.com,ratheesh.m@fmc-asia.com,santhanan.krishnan@fmc-asia.com,anoop.k@fmc-asia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Apheresis Machine <i>Model & Make : COM.TEC/Fresenius Kabi AG</i>	3832640 Incl.GST :12%	282657.2	4115297.2
2	Platelet Kit	7670.88 Incl.GST :12%	565.73	8236.61
3	Plasma Exchange Kit	8678.88 Incl.GST :12%	640.07	9318.95
4	RBC Kit	8678.88 Incl.GST :12%	640.07	9318.95
5	Stem Cell Kit	8678.88 Incl.GST :12%	640.07	9318.95

Item-wise Price Details							
				3866347.52	285143.13	4151490.65	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Apheresis Machine							
Labour	1,02,660.00	1,02,660.00	1,02,660.00	1,02,660.00	1,02,660.00	1,02,660.00	1,02,660.00
Comprehensive	2,05,320.00	2,05,320.00	2,05,320.00	2,05,320.00	2,05,320.00	2,05,320.00	2,05,320.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Apheresis Machine

1. Continuous Flow Technology with single or double needle.
2. Following protocols should be available
 - a. Leucodepleted Platelets + Plasma
 - b. Therapeutic Plasma Exchange
 - c. Peripheral Blood stem cell collection
 - d. Red cell exchange, plasma pheresis
3. Inbuilt process Leucoreduction
4. Extra Corporeal Volume less than 200ml
5. Air Detector
6. Centrifuge leak detector, automatically detected before priming
7. Inlet and return pressure monitor
8. Citrate control and monitoring facility

9. Platelet should be collected with plasma, no manual mixing needed
10. Inbuilt quality control
11. Programmable Haematocrit and Plasma Volume
12. Should have inbuilt battery backup or on-line UPS of suitable capacity with 3 hours backup
13. System should recover from exact point whenever paused
14. 500ml ACD should be provided with plasma exchange / platelet pheresis kits
15. ACD drip/flow monitoring with minimum ACD use
16. Input voltage 230 VAC +10%
17. Haemolysis monitoring for plasma line
18. Yield of 3.3×10^{11} even at pre-count of 2.65 lakhs per micro liter
19. Inbuilt automatic cuff/ Pressure monitoring
20. Onsite training for the departments concerned
21. 20 nos of platelet apheresis kits / plasma exchange kits shall be supplied free of cost as requested by the end user.
22. The rate for the following consumables should be provided in the price bid and the rates shall be freezed for three years from the date of price bid opening. The rate offered for consumables will be taken for L1 evaluation.
 - a. Platelet kit
 - b. Plasma Exchange kit
 - c. RBC Exchange kit
 - d. Stem cell kit

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail