



Running Contract Details	
Equipment Name	Multidirectional Adjustable IV stand (Wall Mounted)
Running Contract Valid Till	17-06-2026
Tender Ref No	KMSCL/EP/T539(R)/365B/2024
Tendered Quantity	3000
Supplier Name	M/s GLOBAL INDUSTRIES
GST No	32EODPK3718A1ZJ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
GLOBAL INDUSTRIES VII/425 KARIMPADAM CHENDAMANGALAM	Contact Person	
	Phone	
	Mobile No	9895331025
	Email	satheesanacs2244@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Multidirectional Adjustable IV stand (Wall Mounted)</b> <i>Model &amp; Make : GL-D005/ Global Industries</i>	3245 Incl.GST :18%	227.15	3472.15
		<b>3245</b>	<b>227.15</b>	<b>3472.15</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Multidirectional Adjustable IV stand (Wall Mounted)**

1. It should be made up of stainless steel (SS304 Grade)
2. It should be fitted on the walls, even on pillars or tiles near patient's cot.
3. The length of IV Stand can be adjusted from 50 cm to 135cm and can be utilized by two patients at a time.
4. It should be shortened and kept closer to wall

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail