



Running Contract Details	
Equipment Name	Portable Green Laser Console with LIO Model B
Running Contract Valid Till	07-01-2023
Tender Ref No	KMSCL/EP/T332/1243B/2019
Tendered Quantity	15
Supplier Name	M/s Nethra Instrument
GST No	33AAFFN1864Q1ZU
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
MIG 134 - 8th Street Mogappair Eri Scheme Chennai - 600037	Contact Person	Mr.R Vasimalai, Mr. R Anandhan
	Phone	044-26530672
	Mobile No	9840033073, 9847063262
	Email	nethra.ints@gmail.com, nethravaasu@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Portable Green Laser Console with LIO Model B <i>Model & Make : Vitra 2/ Quantel medical</i>	1596000 Incl.GST :12%	117705	1713705
		1596000	117705	1713705

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Portable Green Laser Console with LIO Model B							
Labour	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00
Comprehensive	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Portable Green Laser Console with LIO Model B

1. Laser type should be a frequency-doubled solid state laser
2. Should have a 532 nm operating wavelength
3. Should have an aiming laser wavelength of 635 nm and should have independent movements
4. Should have thermoelectric cooling system
5. Should have a output power ranging from 30 mW to 2500 mW
6. Maximum power at the cornea should not exceed 1500 mW
7. Should provide laser indirect ophthalmoscope
8. Should have inbuilt laser safety filter for laser indirect ophthalmoscope
9. Should provide carry case for both LIO and laser console
10. Should provide 20D and 28D lens with laser protective coating
11. Should provide four laser safety goggles for assistant
12. Should have electric requirement of 220V 50 Hz
13. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US). Copy of the certificate/ test report shall be produced along with the technical bid
14. Power supply for LIO should be inbuilt with the console .
15. Should have LED/ halogen illumination and it should be movable.
16. Maximum weight of the whole system (without case/ packing) should be less than or equal to 10 kg