

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/97/2018

Dated: 06.09.2018

**NOTICE**

**Sub:-** Tender No. T271(R)/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Water Bath Large & Binocular Microscope Various Type- reg.

**Ref:-**

1. Tender Inviting Notice published dated 14.02.2018
2. Technical bid opening on 23.03.2018 @ 11.00 A M
3. Clarification notice published on 26.04.2018
4. Demonstration/ document evaluation of Water Bath Large & Binocular Microscope Various Type conducted on 10.07.2018

I. The reports of technical document evaluation/ demonstration of Water Bath Large & Binocular Microscope Various Type as per reference no. 4<sup>th</sup> cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Water Bath Large	M/s. Labline Instruments	SWB-4
		M/s. UNIVERSAL AGENCIES	PLT-109 SEROLOGICAL WATER BATH
2.A	Binocular Microscope Type 1	M/s. Radical Scientific Equipments Pvt. Ltd.	RXL - 5B
		M/s. Scientific Medical Equipments (Kerala) Pvt.	Coax IOB
		M/s. universal agencies	LX 400
3.B	Binocular Microscope Type II	M/s. Radical Scientific Equipments Pvt. Ltd.	RXL - 4
		M/s. universal agencies	LX 500

II. The items offered by the other tenderer participated in the tender are rejected due to the reasons mentioned against their name below.

**III. KMSCL/EP/T265/1169/2018:- Water Bath Large**

A. M/s. Kerala Surgical Equipment Co.

Rejected Model/ OEM: MEDIK-WB20L

Reason for rejection : Sample not Produced

**iv. KMSCL/EP/T265/13/2018 :- Binocular Microscope Various Type**

**1. M/s. AST MEDITECH**

Rejected Model/ OEM: CM - 2100

**A. Binocular Microscope Type -I**

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
7	Illuminator - Built-in LED light source with white light with rechargeable battery backup.	No Battery back up

**B. Binocular Microscope Type -II**

Reason for rejection : Sample Not Produced

**2. M/s. LAWRENCE AND MAYO INDIA PVT LTD**

Rejected Model/ OEM: LM - 52- 1716(N 126)

**A. Binocular Microscope Type -I**

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
3.	Objectives - Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with plan achromatic correction for FOV of 20 mm.	Not, Antifungus coated
7.	Illuminator - Built-in LED light source with white light with rechargeable battery backup.	No Battery back up

**B. Binocular Microscope Type -II**

Reason for rejection : Sample Not Produced

**3. M/s. Scientific Medical Equipments (Kerala) Pvt.**

**B. Binocular Microscope Type -II**

Reason for Rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 07.09.2018 @ 11.00 am at the etender portal.

**Dr. Dileep Kumar S R**  
General Manager

