



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/605/2018

Dated: 05.11.2018

NOTICE

Sub:- Tender No. T275/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening –Various Anesthesia Equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 18.06.2018
 2. Technical bid opening on 16.07.2018 @ 11.00 A M
 3. Eligible List for demonstration notice published on 11.09.2018
 4. Demonstration/ document evaluation of Various Anesthesia Equipments Conducted on 03.11.2018

I. The reports of technical document evaluation/ demonstration of various Anesthesia Equipments as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Multiparameter Monitor	M/s. Sree Gokulam Healthcare Pvt. Ltd.	uMEC 7
		M/s. Skanray Technologies Private Limited	STAR 65
		M/s. Arkon Medical Systems	STAR 8000
		M/s. BPL MEDICAL TECHNOLOGIES PVT LTD	Ultima Prime
2.	Crash Cart	M/s. Midmark (India) Private Limited	9003
		M/s. United Surgical Industries	USI-1070
		M/s. Kerala Surgical Equipment Co	Medik-168
		M/s. CAREVEL MEDICAL SYSTEMS PVT LTD	C-4502
3.	Syringe Pump	M/s. Sree Gokulam Healthcare Pvt. Ltd.	benefusion SP3
		M/s. S AND T Engineers (P) Ltd	HK 400
4.	Infusion Pump	M/s. Sree Gokulam Healthcare Pvt. Ltd.	Benefusion VP3
		M/s. S AND T Engineers (P) Ltd	HK 100II
		M/s. ANAMDEV ENGINEERS	Graseby1200
		M/s. Kerala Surgical Equipment Co.	SN-1600V

II. The equipments offered by the other tenderers participated in the tender are rejected due to the reasons mentioned against their name below.

A. KMSCL/EP/T275/75/2018: Multiparameter Monitor

1. M/s. UNIVERSAL AGENCIES

Rejected Model:MP1200

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7.	Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.	Only 13 parameters

13.	Monitor should have networking facility with bidirectional & bed to bed communication.	Not demonstrated
	Comments	Manufacture name in the CE certificate is M/s Shenzhen where as in the demonstrated monitor it is mentioned as S.CURE and address mentioned as ahmedabad ,India

2. M/s. DEEP MEDITECH PRIVATE LIMITED

Rejected Model : G3D(A3 plus)

Reason for rejection: Sample not produced

3. M/s. Schiller Healthcare India Pvt.Ltd

Rejected Model: Truscope III

Reason for rejection : Sample not produced

4. M/s. AKAS MEDICAL

Rejected Model: Traze 12A

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7	Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.	Only 13 parameters
13.	Monitor should have networking facility with bidirectional & bed to bed communication.	Not demonstrated
11	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.	Detailed Electrical Functional safety Test Report not submitted

5. M/s. Universal Medical Instruments Delhi

Rejected Model: Yk-8000B

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7	Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.	Only 12 parameters
13.	Monitor should have networking facility with bidirectional & bed to bed communication.	Not demonstrated

6. M/s. CURA HEALTHCARE PVT LTD..

Rejected Model: CritiCURA

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7	Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.	Only 13 parameters
13.	Monitor should have networking facility with bidirectional & bed to bed communication.	Not demonstrated

B. KMSCL/EP/T275/286/2018: Crash Cart

1. M/s. WELCRAFT HEALTHCARE PVT. LTD

Rejected Model: EC 165

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Poor workmanship and folding of edges not proper

2. M/s. MEDIMEK INDUSTRIES

Rejected Model: Mi-5028 A

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
10.	Oxygen cylinder stand –Stainless steel, on one side	This feature is not available.

3. M/s. VINOD MEDICAL SYSTEMS PVT LTD

Rejected Model: VWE- 1001

Reason for rejection: Sample not produced.

C. KMSCL/EP/T275/121/2018): Syringe Pump

1. M/s. AKAS MEDICAL

Rejected Model: TERUMAX RX

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
10.	Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.	This feature is not available.

2. M/s. UNIVERSAL AGENCIES

Rejected Model: EN-S7

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Demonstrated different model EN S3

3. M/s. Universal Medical Instruments Delhi

Rejected Model: Dpmed SP1

Reason for rejection:

--	--	--

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No drug library
10.	Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.	This feature is not available

4. M/s. ANAMDEV ENGINEERS

Rejected Model: GRASEBY 2100

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No drug library
10.	Should have automatic bolus reduction system to avoid accidental bolus delivery after occlusion incident.	This feature is not available

5. M/s. CURA HEALTHCARE PVT LTD

Rejected Model: Accupush+

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No dose calculation
	Comments	Display not user friendly and very small font size& difficult to change settings

D. KMSCL/EP/T275/62/2018): Infusion Pump

1 M/s. UNIVERSAL AGENCIES

Rejected Model: IP 201

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No dose mode
	Comments	Manufacture name not available in the demonstrated machine

2. M/s. AKAS MEDICAL

Rejected Model: Infumax

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No dose mode

	Comments	Manufacture name not available in the demonstrated machine
--	----------	--

3. M/s. Universal Medical Instruments Delhi

Rejected Model: DPFUSION VP1

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1.	Should be easy to use and nurse friendly with drug library and dose calculation.	No dose mode

II. The online price bid of the eligible tender mentioned above will be opened on 08.11 .2018 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager