

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-3045647

0471-2337353, 0471-3045600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/1012/2017

Dated: 15.05.2018

**NOTICE**

**Sub:-** Tender No. T261/2017- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Ultrasound with colour doppler portable - reg.

**Ref:-** 1. Tender Inviting Notice published dated 22.12.2017  
2. Technical bid opening on 27.01.2018 @ 11.00 A M  
3. Demonstration/ document evaluation of Ultrasound with colour doppler portable conducted on 14.05.2018

I. The reports of technical document evaluation/ demonstration of Ultrasound with colour doppler portable as per reference no. 3<sup>rd</sup> cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Ultrasound with colour doppler portable	M/s. CURA HEALTHCARE PVT LTD	Vinno V5
		M/s. Cyrix Healthcare Pvt Ltd	Mindray M7 Ultrasound System
		M/s. Esaote India NS Limited	MyLab Gamma

II. The items offered by the other tenderer participated in the tender are rejected due to the reasons mentioned against their name below.

**1.M/s. Gemini Medical Networks Pvt Ltd.**

Rejected Model/ OEM: EBit 60

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
I.2	System should weight not more than 7.5kgs with battery and one single regular probe.	7.8 Kg
I.4	System cold start up time off to on not more than 90 second.	More than 90 seconds
I.11	Should have composite Video output (NTSC/PAL) or HDMI to Video printer or external LCD display.	HDMI port not available
I.16	System should be compatible for TEE/TOE, and should support both in A/C Mains and in battery	Not compatible for TEE probe

III. The online price bid of the eligible tenderer mentioned above will be opened on 17.05.2018 @ 11.00 am at the etender portal.

**Dr. Dileep Kumar S R**  
General Manager