



KERALA MEDICAL SERVICES CORPORATION LTD

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-3045647

0471-2337353, 0471-3045600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

Equipment procurement Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/524/2017

Dated: 08.06.2017

E-TENDER – PRE-TENDER MEETING – T228- AMENDMENT NOTICE

Sub:- e-Tender no. 228/2017 released on 12/05/2017 – amendments published – reg

Ref:- Discussions in the pre-tender meeting held on 24/05/2017

The following amendments are made in respect of the tender clauses noted herein on the basis of the request from prospective tenderers, expert opinion and market study.

TENDER NO. KMSCL / EP / T228 /86A/ 2017 : Operating Microscope ENT Model B

Section No.	Clause No.	Existing clause	Amended clause
Appendix 1	1	Manual 5 step magnification changer with apochromatic optics. Magnification factor 0.4x0.6x -1.0x1.6x - 2.5x integrated green and orange filters	Manual 5 step magnification changer with apochromatic optics and integrated green filters
Appendix 1	4	Pair of wide field push in eyepieces n 10 x with sleeves and magnetic locks, dopeter setting from -8D to +5C	Pair of wide field push in eyepieces10 x / 12x with sleeves and magnetic locks, diopter setting from -5D to +5D
Appendix 1	9	Coaxial fibre optic illumination system 12V 100W	Coaxial fibre optic illumination system
Appendix 1	10	Dual lamp system with quick change over	Dual lamp system with quick change over (Not applicable for LED Light Source)
Appendix 1	16	3CCD camera	HD Camera with HD TV of size 32 inch with C Mount and beam splitter.
Appendix 1	17	Added Clause	Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

Dr. Dileep Kumar S R
General Manager