



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**Equipment Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/600/2019**

**Dated: 08.08.2019**

**NOTICE**

**Sub:-** Tender No. T328(R)/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 07.06.2019
  2. Technical bid opening on 01.07.2019 @ 11.00 am
  3. Clarification Notice published on 11.07.2019 & 23.07.2019
  4. Eligible list for demonstration published on 29.07.2019 & 06.08.2019
  5. Demonstration/ document evaluation of Transcutaneous Bilirubinometer conducted on 27.07.2019
  6. Demonstration/ document evaluation of High Flow Nasal Cannula Therapy Device conducted on 05.08.2019
  7. Demonstration/Document evaluation of Pacemaker Temporary - Single Chamber conducted on 07.08.2019

I. The reports of technical document evaluation/ demonstration of various equipments as per reference no.5th,6th, & 7th<sup>h</sup> cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model/OEM
1.	Transcutaneous Bilirubinometer	M/s Draeger India Pvt Ltd	Draeger JM105/ Dragerwek AG & Co KGaA
2.	High Flow Nasal Cannula Therapy Device	M/s Fisher and Paykel Healthcare India Pvt Ltd	Airvo 2/ Fisher & Paykel Healthcare Corporation Ltd
3.	Pacemaker Temporary - Single Chamber	M/s India Medtronic Pvt Ltd	Medtronic 53401/ Medtronic. Inc.

II. The details of the other tenderers participated in the tender are rejected in the tender due to the reasons mentioned against their name below;

**1. KMSCL/EP/T328(R)/33A/2019 : Pacemaker Temporary - Single Chamber**

A. M/s Sree Gokulumhealth Care

Rejected Model: Pace 101/ Osypka Medical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
-----------	-------------------------	------------------

5.	Should have continuous monitoring of the battery voltage.	This feature is not available.
8.	Should have back up pacing during battery change.	This feature is not available.
12	Should have Pulse Amplitude of 0.1-17V	Only 0.3-12 V

III. The online price bid of the eligible tenderer mentioned above will be opened on 09.08..2019 @ 11.00 am at the e tender portal.

**Dr. Dileep Kumar S R**

കെ.എം.എസ് .സി.എൽ