

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/247/2018

Dated: 06.09.2018

**NOTICE**

**Sub:-** Tender No. T270/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening –Motorized Wheel Chair for Young Paraplegics - reg.

**Ref:-** 1. Tender Inviting Notice published dated 24.03.2018  
2. Technical bid opening on 14.05.2018 @ 11.00 A M  
3. Demonstration / Document evaluation of Motorized Wheel Chair for Young Paraplegics conducted on 12.06.2018

I. The reports of technical document evaluation/ demonstration of Motorized Wheel Chair for Young Paraplegics as per reference no. 3<sup>rd</sup> cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Motorized Wheel Chair for Young Paraplegics	M/s. Karma Healthcare Limited	KM 25.2
		M/s. New India Surgicals & Chemicals	2974

II. The items offered by the other tenderer participated in the tender are rejected due to the reasons mentioned against their name below.

1. **M/s.** Ostrich Mobility Instruments Pvt. Ltd

Rejected Model/ OEM: Verve Fx

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
2.	Should have Rear wheels> 25 cm solid	No,Only 16 cm

**2. M/s. New Glory Orthopaedics**

Rejected Model/ OEM: MWC - 0518

Reason for rejection : Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 11.09.2018 @ 11.00 am at the etender portal.

**Dr. Dileep Kumar S R**  
General Manager