

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/629/2018

Dated: 08.10.2018

**NOTICE**

**Sub:-** Tender No. T282(R)/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening –various equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 18.06.2018
  2. Technical bid opening on 11.07.2018 & 01.08.2018@ 11.00 A M
  3. Eligible List for Equipment Demonstration published on 11.09.2018
  4. Report of Demonstration/ document evaluation of 3D Laproscopic set with Instruments received on 04.10.2018
  5. Report of Demonstration/ document evaluation of Automated Blood Culture System Model B & Medicine Rack received on 06.10.2018

I. The reports of technical document evaluation/ demonstration of various equipments as per reference no. 4<sup>th</sup> & 5<sup>th</sup> cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	3D Laproscopic set with Instruments	M/s. Lakshmi Surgicals	Karl Storz GmbH & Co Germany
2.	Automated Blood Culture System Model B	M/s. ABC DIAGNOSTICS PVT. LTD	Vitek 30
3.	Medicine Rack	M/s. Shuter Enterprises India Pvt Limited	Shuter

II. The items offered by the other tenderer participated in the tender are rejected due to the reasons mentioned against their name below.

A. KMSCL/EP/T282(R)/1131/2018 : 3D Laproscopic set with Instruments

1.M/s. Olympus Medical Systems India Pvt. Ltd

Rejected Model/ OEM : Olympus Corporation

Reason for rejection :

Clause No	Technical Specification	Deviation Noted
	Should also have a facility to connect routine camera	

10	heads for general surgery and urology for use with non video telescopes.	This feature is not available.
12	Should have facility for automatic brightness control for best distribution of light evenly over the image field	This feature is not available.
13	Should have facility for capturing & storing 3D/2D Images & videos in USB provided with system.	This feature is not available.
Comments :		Depth perception is not satisfactory. Image quality is not satisfactory. Doesnot match with specification.

III. The online price bid of the eligible tenderer mentioned above will be opened on 10.10.2018 @ 11.00 am at the etender portal.

**Dr. Dileep Kumar S R**  
General Manager