



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/606/2018

Dated: 08.10.2018

NOTICE

Sub:- Tender No. T276/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening –Various Equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 18.06.2018
 2. Technical bid opening on 13.08.2018@ 11.00 A M
 3. Eligible List for Equipment Demonstration published on 11.09.2018
 4. Demonstration/ Document evaluation of various equipments conducted on 08.10.2018

I. The reports of technical document evaluation of various equipments as per reference no. 4th cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	OT Table Manual Hydraulic	M/s. Kerala Surgical Equipment Co.	PSI -Virtue
		M/s. Trichur Surgicals	TS 105
2.	Diathermy Surgical	M/s. JBS Mediitec India P Ltd	ELSY 360 L+
		M/s. India Medtronic Pvt. Ltd.	Wem SS-501 SX
		M/s. Choice Medisystems	SIGMA+

II The items offered by the other tenderers participated in the tender are rejected due to the reasons mentioned against their name below.

A. KMSCL/EP/T276/92/2018 : OT Table Manual Hydraulic

1. M/s. United Surgical Industries

Rejected Model : USI-2001

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
4.	The table top should have a minimum vertical stroke of 250mm.	At the time of initial demonstration height shown only upto 960 mm later demonstrated upto 1030 mm which is not satisfactory.
16.c	A pair of padded shoulder support with clamps - 1	Adjustment not possible

16.d	A pair of padded lateral support with clamps - 1	Adjustment not possible
	Comments	1. The internal mechanism for movements is exposed, which is difficult to clean and leads to infection 2. Table top Shaking 3. Over-all quality is not satisfactory

III. Due to technical reason the tender for : KMSCL/EP/T276/90/2018: OT Light Shadowless Ceiling Mount Single Dome is cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on 10.10.2018 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager