

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/97/2018

Dated: 04.09.2018

NOTICE

Sub:- Tender No. T265/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Hematology Analyser-3part- reg.

Ref:- 1. Tender Inviting Notice published dated 14.02.2018
2. Technical bid opening on 23.03.2018 @ 11.00 A M
3. Demonstration/ document evaluation of Hematology Analyser-3part conducted on 16.07.2018 & 17.07.2018

I. The reports of technical document evaluation/ demonstration of Hematology Analyser-3part as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Hematology Analyser-3part	M/s. AGD BIOMEDICALS PVT LTD	ERMA PCE – 210
		M/s. HORIBA India Private Ltd	HHORIBA/ABX MICROS ES 60
		M/s. UNIVERSAL AGENCIES	SMART- I
		M/s. VECTOR BIOTEK PVT.LTD	URIT 3000 PLUS

II. The following items offered by the tenderes participated in the tender are rejected due to the reason mentioned against their name below ;

1. M/s. Agappe Diagnostics Ltd,

Rejected Model/ OEM: MISPA COUNT

Clause No	Technical Specification	Deviations Noted
13	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%

2 .M/s. DIATEK HEALTHCARE PVT LTD.

Rejected Model/ OEM: RT 7600 S

Clause No	Technical Specification	Deviations Noted
13	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%

3 .M/s. insignia biomedical

Rejected Model/ OEM: Hemax 330

Clause No	Technical Specification	Deviations Noted
	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment.	

13	The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%
----	---	-----------------

4.M/s. LabX

Rejected Model/ OEM: Medonic M32 B

Clause No	Technical Specification	Deviations Noted
13	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%

5 .M/s. Meril Diagnostics Pvt Ltd

Rejected Model/ OEM: Merilyzer celQuant 360

Clause No	Technical Specification	Deviations Noted
13	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%

6. M/s.Transasia Bio Medicals Ltd

Rejected Model/ OEM: Sysmex -Xp 100

Clause No	Technical Specification	Deviations Noted
13	3 level QC should be demonstrated to the technical committee during the technical demonstration of the offered equipment. The sequence of QC run (all 3 levels) will be intimated during demo. The CV of RBC,WBC,PLT and HB should be less than or equal to 5%.	No,More than 5%

III. The online price bid of the eligible tenderer mentioned above will be opened on 06.09 .2018 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager