



KERALA MEDICAL SERVICES CORPORATION LTD

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-3045647

0471-2337353, 0471-3045600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

Equipment procurement Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/524/2017

Dated: 08.06.2017

E-TENDER – PRE-TENDER MEETING – T228- AMENDMENT NOTICE

Sub:- e-Tender no. 228/2017 released on 12/05/2017 – amendments published – reg

Ref:- Discussions in the pre-tender meeting held on 24/05/2017

The following amendments are made in respect of the tender clauses noted herein on the basis of the request from prospective tenderers, expert opinion and market study.

TENDER NO. KMSCL / EP / T228 / 58 / 2017 : ICU COT Model B

Section No.	Clause No.	Existing clause	Amended clause
Appendix 1	1	Overall dimensions should have (L) 2270 mm X (W) 1050 mm X (H) Adjustable from 470mm to 760mm (approximate, minor variation will be accepted)	Overall dimensions should have (L) 2270 mm X (W) 1050 mm X (H) Adjustable from 470mm to 760mm (approximate, minor variation will be accepted, ±5%)
	3	It should have PPE Head & Foot board fixed on the bed frame with an inbuilt nylon sleeve; 4 PPE side boards with 360 deg swivel mechanism and push button mechanism for operation	It should have PPE Head & Foot board fixed/ detachable on the bed frame with an inbuilt nylon sleeve; 4 PPE side boards with 360 deg swivel mechanism and push button mechanism for operation
	4	It should have 4 section lying surface, 3 made of LDPE having Bacteriostatic additives & Joined by corrosion-resistant joineries. The bacteriostatic additives are formulated to fulfill the requirements for bacterial protection against at least 2 commonly found bacteria in Hospital environment)	It should have 4 section lying surface, 3 made of LDPE/PPE/ABS having Bacteriostatic additives & Joined by corrosion-resistant joineries. The bacteriostatic additives are formulated to fulfill the requirements for bacterial protection against at least 2 commonly found bacteria in Hospital environment)
	13	It should have Min ht: 460 mm; Max ht:760 mm (As per NABH norms) (All dimensions approximate only, minor variations acceptable)	It should have Min ht: 460 mm; Max ht:760 mm (As per NABH norms) (All dimensions approximate only, minor variations acceptable, ±5%)
	17	The bed should adhere to EN norms for design & USFDA norms for patient safety.	The bed should adhere to EN 60601-2-52 for design & USFDA norms for patient safety.
	23	Added Clause	Should provide with removable heavy duty saline stand made of stainless steel.304 grade and one telescopic saline stand with two hooks. . Should have a provision to fix the

			telescopic saline stand on all four corners of the bed and also in the middle of the bed on either side
--	--	--	---

Dr. Dileep Kumar S R
General Manager