



KERALA MEDICAL SERVICES CORPORATION LTD

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-3045647

0471-2337353, 0471-3045600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

Equipment procurement Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/524/2017

Dated: 08.06.2017

E-TENDER – PRE-TENDER MEETING – T228- AMENDMENT NOTICE

Sub:- e-Tender no. 228/2017 released on 12/05/2017 – amendments published – reg.

Ref:- Discussions in the pre-tender meeting held on 24/05/2017

The following amendments are made in respect of the tender clauses noted herein on the basis of the request from prospective tenderers, expert opinion and market study.

TENDER NO. KMSCL / EP / T228 /58C/ 2017 : ICU COT Model C

Section No.	Clause No.	Existing clause	Amended clause
Appendix 1	8	Should provide with a collapsible split side board made up of ABS/PPE. It should be easy to operate and have damping mechanism. Side boards shall go below the bed height and should not interfere the patient coming out of bed	Should provide with a swing away split side board made up of ABS/PPE. It should be easy to operate and have damping mechanism. Side boards shall go below the bed height and should not interfere the patient coming out of bed
	10	Removable mattress platform of suitable size made up of ABS/SS304 should be provided	Removable mattress platform of suitable size made up of ABS/PPE should be provided
	13	Height range of 38cm to 79 cm (+/- 7cm)	Height range of 38cm to 79 cm (+/-7cm) without mattresses.
	19	Lockout switches on attendant Control Panel	Lockout switches on attendant Control Panel, single press chair position and emergency head down.
	30	Added Clause	Antimicrobial powder Coat
	31	Added Clause	Should provide with removable heavy duty saline stand made of stainless steel.304 grade and one telescopic saline stand with two hooks. . Should have a provision to fix the telescopic saline stand on all four corners of the bed and also in the middle of the bed on either side

Dr. Dileep Kumar S R
General Manager