



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**Equipment Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/323/2019**

**Dated: 03.09.2019**

**NOTICE**

**Sub:-** Tender No. T323/2019- Results of representation technical document evaluation/ demonstration & Date of Price Bid Opening – Arthroscopy Camera and Instruments and Shaver System Model B- reg.

- Ref:-**
1. Tender Inviting Notice published dated 01.04.2019
  2. Technical bid opening on 07.04.2019, 14.05.2019 & 21.05.2019@ 11.00 A M
  3. Clarification notice published on 15.05.2019 & 22.05.2019.
  4. Eligible List for demonstration published on : 06.06.2019
  5. Demonstration/ document evaluation of Arthroscopy Camera and Instruments and Shaver System Model B conducted on 08.07.2019 .
  6. Result of demonstration published on 19.08.2019.
  7. Representation email received on 19.08.2019
  8. Reports of Evaluation of various representations and their clarifications of Arthroscopy Camera and Instruments and Shaver System Model B received on 03.09.2019

I. With reference to the demonstration result notice published for Arthroscopy Camera and Instruments and Shaver System Model B, representations were submitted by tenderers against the result notice. The representation received was evaluated in detail and the findings and decisions are as follows;

<b>The complaint received against the selection of the offered model (Dyonics 2/Smith and Nephew) of M/s Axis Medical Systems.Ltd</b>			
<b>Clause No</b>	<b>Complaint</b>	<b>Clarification Submitted</b>	<b>Findings of the committee</b>
	Axis have quoted Smith & Nephew Lens Camera system. This is single chip camera. We are attaching the official statement from Smith &	<ol style="list-style-type: none"> <li>1. Smith and Nephew Lens system work in the latest semiconductor based technology ( CMOS)</li> <li>2. It's HD with 1920X1080P output.</li> <li>3. System offers future all upgrades including 4K upgrade in the same platform.</li> </ol>	Technical committee verified the complaint and the clarification submitted by the bidder. M/s Axis Medical Systems Ltd offered model was a single chip HD camera whereas the stipulated

No.II.13	Nephew to US FDA which clarifies this. This document is publically available on the US FDA website.	<p>4. How a single chip can be upgraded to 4K technology</p> <p>5. System is been installed more than 100 units across the country in all very important hospitals like SRMC, Fortis, Ortho one, Cosmo, Meitra, Apollo and many more reputed hospitals.</p>	<p>specifications is for 3 chip HD camera. Hence the technical committee recommended to reject the offer of M/s Axis Medical Systems Ltd</p>
<p><b>Recommendations of the committee:</b> The complaint received against M/s Axis Medical Systems is considered for further tender process.</p>			

II.The details of the tenderers qualified after the evaluation for price bid opening is tabulated as follows

SI. No	Equipment	Name of the Selected Bidders	Selected Model/OEM
1.	Arthroscopy Camera & Instruments and Shaver System Model B	M/s Karl Storz	Arthroscopy Camera and Instruments and Shaver System, Karl Storz SE & Co.KG
		M/s Neoptic Surgicals	1588 AIM Camera Control Unit/ Stryker

III.The online price bid of the eligible tenderers mentioned above will be opened on 06. 09.2019 @ 11.00 am at the e tender portal.

**Dr. Dileep Kumar S R**

കെ.എം.എസ് .സി.എൽ