



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**Equipment Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/600/2019**

**Dated: 07.09.2019**

**NOTICE**

**Sub:-** Tender No. T328(R)/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments - reg.

- Ref:-**
1. Tender Inviting Notice published dated 07.06.2019
  2. Technical bid opening on 01.07.2019 @ 11.00 am
  3. Clarification Notice published on 11.07.2019
  4. Eligible list for demonstration published on 29.07.2019
  5. Reports of demonstration/ document evaluation of Shower Trolley received on 06.09.2019
  6. Demonstration/ document evaluation of Shower Trolley conducted on 06.09.2019

I. The reports of technical document evaluation/ demonstration of various equipments as per reference no.5th & 6th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model/OEM
1.	OT Table Orthopedics Manual Hydraulic	M/s Kerala Surgical Equipment Co	PSI -Virtue/Palakkad Surgicals
2.	Shower Trolley	M/s United Surgical Industries	USI-1077/United Surgical Industries

II. The details of the other tenderers participated in the tender are rejected in the tender due to the reasons mentioned against their name below;

**1. KMSCL/EP/T328(R)/93/2019 : OT Table Orthopedics Manual Hydraulic**

**A. M/s Trichur Surgicals**

Rejected Model: TS 110/Trichur Surgicals

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	The pelvic support should be movable to the	Pelvic support initially demonstrated was not movable to the edges later they

17.	edges and completely radiolucent.	demonstrated another pelvic support which was very loose when fixed with table
-----	-----------------------------------	--

**B.M/s United Surgical Industries**

Rejected Model; USI-2003 + 2006/United Surgical Industries

Reason for rejection: Sample not produced

**C. M/s Confident Dental Equipments Ltd**

Rejected Model: MHOTO14/Confident

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
16	The table should be compatible for spine surgeries with C-arm or separate accessory should be provided for doing spine surgery with C-arm.	The position demonstrated with spine accessory is not c-arm compatible.
19.c	A pair of padded shoulder support with clamps - 1	Not demonstrated
19.d	A pair of padded lateral support with clamps - 1	Not demonstrated
19.e	A pair of leg crutches with clamps - 1	Not demonstrated
19.g	Adjustable traction device with minimum leg stroke of 175mm	Minimum leg stroke only 150 mm

**2. KMSCL/EP/T328(R)/1350/2019 : Shower Trolley**

**A. M/s Kerala Surgical Equipment Co**

Rejected Model: MDK-548/Medik Enterprises

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 16.09..2019 @ 11.00 am at the e tender portal.

**Dr. Dileep Kumar S R**

കെ.എം.എസ് .സി.എൽ

